

DRY BATH INCUBATOR BAC19-1N



DRY BATH INCUBATOR BAC19-1N

Compact design with stable operation, easy cleaning, replacement and disinfection of metal blocks. It is widely used for enzyme reactions, immunoassays, melting and boiling points, incubation and activation of cultures and variation of laboratory procedure. Flexible to accommodate a variety of interchangeable heating blocks.

Used in Clinical, General Chemistry, Preservation, Reactions of sample, DNA amplification, Initial Denaturation of electrophoresis.

BAC19-1N DRY BATH INCUBATOR

Fast heating speed, uniform heating, accurate temperature control, high stability, low energy consumption and no noise.

Simple or program temperature control mode is available, making the experiment more convenient.

Sliding operation, light touch button, novel and fashionable.

The standard metal blocks are easy to replace, easy to clean and disinfect.

Protect samples from contamination.

The buzzer tone can be turned off, making the experiment quieter.



SPECIFICATIONS

Model	BAC19-1N
Temperature Range	RT+5- 150°C
Temp. Setting Range	5°C~150°C
Time Range	1min~99h59m (00:00 is continuous)
Block Temp. Stability@40~100°C	±0.5 °C
Block Temp. Stability@> 100°C	±1 °C
Block Temp. Uniformity@40°C	±0.3 °C
Block Temp. Uniformity@>40°C	±0.5 °C
Temp. Display Accuracy	0.1 °C
Heating Speed	≤15min(20°C~150°C)
Sample Capacity	1 standard block
Voltage	AC 220V/AC 110V 50Hz/60Hz
Power	200W
Fuse	250V,2A/3A,φ5x20
Dimension	W.285xD.225xH.95
Net Weight(kgs)	2.48 kgs

OPTIONAL ACCESSORIES

Accessory Code	Name	Description	Capacity
1702028005	Block A	6 mm	42
1702028006	Block B	7 mm	42
1702028007	Block C	10 mm	20

1702028008	Block D	12 mm	20
1702028009	Block E	13 mm	20
1702028010	Block F	15 mm	12
1702028011	Block G	16 mm	12
1702028012	Block H	19 mm	12
1702028013	Block I	20 mm	6
1702028014	Block J	26 mm	6
1702028015	Block K	28 mm	4
1702028016	Block L	40 mm	2
1702028017	Block M	0.5 mm	42
1702028018	Block N	1.5 mm	24
1702028019	Block O	2.0 mm	24
1702028020	Block P	0.2 mm	48

LABSTAC

Labstac LLC

82 Wendell Avenue, STE 100, Pittsfield, MA, 01201, USA

Email: contact@labstac.com | Website: labstac.com