

CONDUCTIVITY METER MET31-200



CONDUCTIVITY METER MET31-200

Conductivity meter is an instrument that measures the amount of electric conductivity or current in a solution. The conductance can be measured by applying an alternating electrical current to the two electrodes present in the solution, after which the cations move to the negative electrode and the anions move to the positive electrode. This movement ultimately leads the solution to be conductive.

MET31-200 CONDUCTIVITY METER



SPECIFICATIONS

Model	MET31-200
Conductivity	
Range	0.00 μS/cm to 200 mS/cm
Resolution	1.01 μS/cm minimum; changed with range
Accuracy	± 1.0 % FS
Reference Temperature	25 °C
Standard Recognition	84 μS/cm, 1413 μS/cm; 12.88 mS/cm
TDS	
Range	0.00 mg/L ~100 g/L
Resolution	0.01mg/L minimum; changed with range
Accuracy	± 1.0% FS
Temperature	
Range	- 5 to 110 °C, 23 to 230 °F
Unit	°C,
Resolution	0.1
Accuracy	±0.2
Measurement	
Reading Mode	AutoRead, Continuous
Reading Prompts	Reading, Stable, Locked
Temp. Compensation	ATC, MTC
Data Management	
Data Storage	50 results each
Inputs	
Temp./EC. Probe	5-pin aviation connector

2

Yes
300, 600, 1200, 1800, 3600sec, off
IP54
AC Adapter,100-240 V AC
242 x 195 x 68 mm
900 g(1.98 lb)



Labstac LLC

82 Wendell Avenue, STE 100, Pittsfield, MA, 01201, USA Email: contact@mail.labstac.com | Website: labstac.com