

LARGE CAPACITY DEIONIZED WATER SYSTEM WPS16-090



LARGE CAPACITY DEIONIZED WATER SYSTEM

WPS16-090

It provides a variety of applications from residential to scientific and industrial settings. It completely meets the requirements of general chemical or biological experiments for pure water. Deionized water system is an ideal choice of deionized water for grade experiments.

Used in Laboratory, Manufacturing, Reefkeeping, Aquarium.

Also known as Laboratory Deionized water system.

WPS16-090 LARGE CAPACITY DEIONIZED WATER SYSTEM

Integration design

Integrating pretreatment, reverse osmosis, deionization, ultraviolet, ultrafiltration, microfiltration, 250 liters stainless steel tank and pure water supplying and circulation system together.

Perfect control, monitor and alarm

This series could monitor and alarm, including shortage of water, leaking, water pressure, water level, flow velocity and water quality etc.

Operate and record easily

This series operate automatically, all the status of working has indicator light; it also could connect to the computer, then you can download all the information from the computer.

Reliable safety

This series would alarm, when the water quality is not qualified, also has the protection of high/low voltage, electrical overload protection and protection for leaking.

Good extension

BCPS 600 series pure water could be feed water of BBPS 200, BDPS 400, BLPS 100, BLPS 200 and BLPS 300 series. The quality of ultrapure water can reach to 18.2MΩ.cm, meet the requirements of PLC, IC, ICP-MS, GF-AAS、Physics, electrochemical and interface research, molecular biology and life science, animal cells and plant cell culture.



SPECIFICATIONS

Model	WPS16-090
Feed Water Requirements*	
Water Inlet	Tap water
Temperature	5-45°C
Pressure	1.0-4.0 Kg/cm ²
Bacteria	<0.1 cfu/ml
DimensionLxWxH	570x600x1500 mm
Weight	60 kg
Power Consumption (W)	240 W
Power Supply	AC110-220 V, 50/60 Hz
Note	*The quality of output water accords with the quality of inlet water.
Deionized water quality	
Resistivity	> 10 MΩ.cm

Conductivity	-
Particle(>0.2µm)	-
Ultrapure Water Quality	
Heavy metal ion	<0.1 ppb
Feed Water Requirements	
Output	90 L/hrs
Conductivity of RO water quality	< tap waterx4%

LABSTAC

Labstac LLC

82 Wendell Avenue, STE 100, Pittsfield, MA, 01201, USA

Email: contact@labstac.com | Website: labstac.com